

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 03 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130295</u>	
Name <u>Joe Corbett</u>		ID# <u>4971</u>	Date <u>5/19/10</u>
Agency <u>Bethel Police Dept</u>		Phone # <u>545-0139</u>	
Instrument Location <u>Bethel, AK Police Dept</u>			
Alco S/N <u>X301563</u>		Target Value <u>0.079</u> <u>0.081</u> <sup>6/7/10</sup>	High Pressure <u>600 psi</u>
Alco Test Values		1 <sup>st</sup> Alco <u>0.079</u> <u>0.081</u> <sup>6-3-10</sup> <sub>CMN</sub>	2 <sup>nd</sup> Alco <u>0.081</u> <u>0.080</u> <sup>6-3-10</sup> <sub>CMN</sub>
Signature <u>[Signature]</u>		<u>CB</u> <u>6/7/10</u>	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 21<sup>st</sup> day of June, 2010.

[Signature]  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130295

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130295

MAY 19, 2010

OPERATOR'S NAME:

CORBETT/JOE

OPERATOR'S NUMBER: 4971

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 123456

DEPT/AGENCY: BET1

CASE/REPORT: UOC 5/19/2010

TEST TYPE: T

ALCO TARGET VALUE: .079

ALCO S/N: X301563

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.96 in  
ALCO TARGET .079 04:57  
BLANK TEST .000 04:58  
INTERNAL STANDARD VERIFIED 04:58  
ALCO TV 29.96 in .001 04:58  
BLANK TEST .000 04:59  
SUBJECT SAMPLE .000 05:01  
BLANK TEST .000 05:01  
ALCO TV 29.96 in .000 05:02  
BLANK TEST .000 05:02

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130295

MAY 19, 2010

TIME 04:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 39c

BAROMETER: 29.98 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~